

Customer Site:
 Peak Waste
 Wood Lane
 Kniveton
 Ashbourne
 DE6 1JF

Certificate of Accuracy 3601
 Date: 4-3-20

Platform Type : 2SF	No. of load points : 6
Indicator Type : DFWT	
Serial No. : 0100126256	
Maximum Cap. : 50000kg	Minimum Capacity (20 x e) : 400kg
Increment (e) : 20kg	Class : III

ECCENTRIC LOADING
 Accuracy at load points to be tested:
 Max. Capacity = Nearest
 No./load points - 1 practical weight

50000 = 10000 KG
 (6-1) Pass Fail

Accuracy across load points (Expanded Display)

<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>

Weigh Office

Accuracy of Instrument - Linearity & Hysteresis (Expanded display)

Known wt.	Increasing reading	Decreasing reading	Error
400	400	400	0
11400	11400	11400	0
22400	22400	22400	0
33800	33800	33800	0
43800	43800		0

Pass
 Fail

ROLLING TESTS LOAD 43800 kg (Expanded display)

END	MIDDLE	END	Pass <input checked="" type="checkbox"/>	END	MIDDLE	END
43800	43800	43800	Fail <input type="checkbox"/>	43800	43800	43800

Deviations recorded as fractions of (e) unless expanded display used and recorded in kg.

	TEST	EC Verification
Permissible errors:	0 ≤ m ≤ 500e	+/-1e
	500 ≤ m ≤ 2000e	+/-2e
	2000 ≤ m ≤ 10000e	+/-3e
		+/-0.5e
		+/-1e
		+/-1.5e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

Undertaken by LC Burgess

No. TR10658 dated 12-3-19