



FS 57316

Customer Site:
 Peak Waste Ltd.
 Wood Lane
 Kniveton
 Ashbourne DE613F

Certificate of Accuracy

3491

Date: 22-5-19

Platform Type : 2SF-30	No. of load points : 6
Indicator Type : DFWT	
Serial No. : 0100126256	
Maximum Cap. : 50000kg	Minimum Capacity (20 x e) : 4000kg
Increment (e) : 20kg	Class : III

ECCENTRIC LOADING

Accuracy at load points to be tested:
 Max. Capacity = Nearest
 No./load points - 1 practical weight

50000 = 10000 KG
 (6-1)

Pass Fail

Accuracy across load points (Expanded Display)

✓	✓	✓
✓	✓	✓

Weigh Office

Accuracy of Instrument - Linearity & Hysteresis (Expanded display)

Known wt.	Increasing reading	Decreasing reading	Error
1000	1000	1000	0
11000	10980	10980	-1e
22000	21960	21960	-2e
33700	33660	33660	-2e
43700	43640	43640	-3e

Pass

Fail

ROLLING TESTS LOAD 43700 kg (Expanded display)

END	MIDDLE	END	Pass <input checked="" type="checkbox"/>	END	MIDDLE	END
43640	43640	43640	Fail <input type="checkbox"/>	43640	43640	43640

Deviations recorded as fractions of (e) unless expanded display used and recorded in kg.

	TEST	EC Verification
Permissible errors:	0 ≤ m ≤ 500e	+/- 0.5e
	500 ≤ m ≤ 2000e	+/- 1e
	2000 ≤ m ≤ 10000e	+/- 1.5e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

Undertaken by IC Burgess

No. TR10658 dated 12-3-19