



Peak Waste Ltd  
Wood Lane  
Kniveton  
Ashbourne  
DE6 1JF

## Certificate of Accuracy

3793

Date: 26-3-21

Platform Type : 25FA-30	No. of load points : 6
Indicator Type : DFWL	
Serial No. : 0100126256	
Maximum Cap. : 50000kg	Minimum Capacity (20 x e) : 4000kg
Increment (e) : 20kg	Class : III

### ECCENTRIC LOADING

Accuracy at load points to be tested:

Max. Capacity = Nearest  
No./load points - 1 practical weight

50000 = 10000 KG

(6-1)

Pass  Fail

Accuracy across load points (Expanded Display )

✓	✓	✓
✓	✓	✓

Weigh Office

Accuracy of Instrument - Linearity & Hysteresis (Expanded display )

Known wt.	Increasing reading	Decreasing reading	Error
400	400	400	0
11400	11400	11400	0
22400	22400	22400	0
33860	33840	33840	-20
43900	43880		-20

Pass

Fail

ROLLING TESTS LOAD... 43880 kg (Expanded display )

END	MIDDLE	END	Pass <input checked="" type="checkbox"/>	END	MIDDLE	END
43880	43880	43880	Fail <input type="checkbox"/>	43880	43880	43880

Deviations recorded as fractions of (e) unless expanded display used and recorded in kg.

	TEST	EC Verification
Permissible errors:	0 ≤ m ≤ 500e	+/- 0.5e
	500 ≤ m ≤ 2000e	+/- 1e
	2000 ≤ m ≤ 10000e	+/- 1.5e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

Undertaken by

A Gcett

No. TR11209 dated 23-3-21