



Peak Waste Ltd  
Wood Lane  
Kniveton  
Ashbourne  
DE6 1JF

## Certificate of Accuracy

4337

Date: 11-4-24

Platform Type : 15m 5155	No. of load points : 6
Indicator Type : 3590	
Serial No. : 14513075	
Maximum Cap. : 50000 kg	Minimum Capacity (20 x e) : 400 kg
Increment (e) : 20 kg	Class : III

**ECCENTRIC LOADING**  
Accuracy at load points to be tested:  
Max. Capacity = Nearest  
No./load points - 1 practical weight

50000 = 10000 KG  
(6-1) Pass  Fail

Accuracy across load points (Expanded Display )

✓	✓	✓
✓	✓	✓

Weigh Office

Accuracy of Instrument - Linearity & Hysteresis (Expanded display )

Known wt.	Increasing reading	Decreasing reading	Error
400	400	400	0
11400	11400	11400	0
22400	22400	22400	0
33920	33900	33900	-20
43920	43900		-20

Pass   
Fail

**ROLLING TESTS** LOAD 43900 kg (Expanded display )

END	MIDDLE	END	Pass <input checked="" type="checkbox"/>	END	MIDDLE	END
43900	43900	43900	Fail <input type="checkbox"/>	43900	43900	43900

Deviations recorded as fractions of (e) unless expanded display used and recorded in kg.

	TEST	Verification
Permissible errors:	0 ≤ m ≤ 500e	+/- 1e
	500 ≤ m ≤ 2000e	+/- 0.5e
	2000 ≤ m ≤ 10000e	+/- 1e
		+/- 1.5e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

Undertaken by N Burgess

No. TR 12336 dated 14-3-24